

Certified Reference Material

Certificate of Analysis

Product ID: MBH-55X G02D9 L

ISO
17034:2016

ISO/IEC
17025:2017

ISO
9001:2015

Product Description: Al/Si/Cu Alloy

Revision No.: 001
Revision Date: 04/06/2022

Description and Intended Use: This Certified Reference Material is covered under the scope of accreditation to ISO 17034 by LGC Standards - Manchester, NH. As an ISO 17034 certified reference material, appropriate use of this material will fulfill the certified reference material and traceability requirements for use in ISO 17025 accredited laboratories. This CRM may come in the form of a solid disk, or chips. The intended use of this CRM may include, but is not limited to, the calibration of instruments and the validation of analytical methods.

Certified Values listed in wt.% with associated uncertainties								
Ag	0.0005	±0.0002	Be	0.00029	±0.00006	Bi	0.026	±0.002
Co	0.0054	±0.0003	Cr	0.086	±0.004	Cu	3.44	±0.04
Ga	0.013	±0.001	Mg	0.12	±0.01	Mn	0.200	±0.003
P	0.003	±0.001	Pb	0.134	±0.006	Sb	0.051	±0.003
Sn	0.27	±0.01	Ti	0.105	±0.004	V	0.0171	±0.0005
Zr	0.0010	±0.0003						

Indicative Values listed in ppm

Al (84.4%) Na (200) S (60) Sr (10)

Homogeneity and Uncertainty: "Uncertainty" values, as reported adjacent to certified concentration values, are based on a 95% Confidence Interval. These estimated uncertainties include the combined effects of method imprecision, material inhomogeneity, and any bias between methods. Homogeneity data from experimental XRF results are reflected in both the overall statistics and certified data. Homogeneity samples are selected by a systematic sampling procedure. The number of samples may be determined by equation 1, where N_{prod} is the number of units produced and N_{min} is the number of samples used for homogeneity testing. These samples are arranged in a simple randomized design such that each sample is analyzed multiple times by XRF. Homogeneity may also be determined within sample using an applied version of ASTM E826. A single factor ANOVA is used to calculate uncertainty due to inhomogeneity (U_{hom}). Uncertainty of the material is calculated by equation 2, where $H=U_{hom}$, S = Standard deviation, t = t-value at 95% CI, and n = number of observations.

$$1. N_{MIN} = \max(10, \sqrt[3]{N_{PROD}})$$

$$2. U_{CRM} = \frac{\sqrt{H^2 + S^2}}{\sqrt{n}} * t$$

Certification Laboratories: Much of the analytical work performed to assess this material has been carried out by laboratories with proven competence, as indicated by their accreditation to ISO 17025. It is an implicit requirement for this accreditation that analytical work should be performed with due traceability, via an unbroken chain of comparisons, each with stated uncertainty, to primary standards such as the mole, or to nationally- or internationally-recognised reference materials. Of the individual results herein, some have traceability (to the mole) via primary analytical methods. Some are traceable to substances of known stoichiometry. Most have traceability via commercial solutions. Furthermore, some results have additional traceability to NIST standards, as part of the analytical calibration or process control.

- LGC Standards - Manchester, NH
- New Hampshire Materials Laboratory - Somersworth, NH
- Applied Technical Services - Marietta, GA
- Lucid Laboratories - Telangana, India
- Scrooby's Laboratory Services - Benoni, South Africa
- Dirats Laboratories - Westfield, MA
- NSL Analytical Services - Cleveland, OH
- IMR Test Labs - Louisville, KY
- Genitest Inc. - Montreal, Canada
- SGS MSI - Melrose Park, IL
- TEC Eurolab - Campogalliano, Italy
- Laboratory Testing, Inc. - Hatfield, PA
- AnchorCert Analytical - Birmingham, UK
- Universal Scientific Laboratory - Revesby, Australia

Instructions for Use: The test surface is on the opposite side of the labeled surface, which includes the material identification. The entire thickness of the unit is certified. However, the user is cautioned not to measure disks less than 2 mm thick when using X-ray fluorescence spectrometry. Each packaged disk has been prepared by finishing the test surface using a lathe. The user must determine the correct surface preparation procedure for each analytical technique. The user is cautioned to use care when either resurfacing the disk or performing additional polishing, as these processes may contaminate the surface. The minimum sample size for chips should be individually evaluated based on the analytical technique used; this would typically be greater than 0.1 grams. The material should be stored in a cool, dry location when not in use.

Chips are not recommended for gas analysis.

Period of Validity: The certification of this material is valid indefinitely, within the uncertainty specified, provided the material is handled and stored in accordance with the instructions stated on this certificate. The certification is nullified if the material is damaged, contaminated, otherwise modified, or used in a manner for which it was not intended.



Kimberly Halliotis, Global Product Manager

April 06, 2022
Certification Date



ISO 17034 Accredited: Reference Materials
Producer, Certificate # 2848.02
ISO/IEC 17025 Accredited: Chemical
Testing, Certificate # 2848.01

Conditions of Sale and Supply: All CRMs & RMs sold are subject to applicable LGC Standard Terms and Conditions of Sale.



The following data represents all pertinent information reported as it applies to the chemical characterization of this material.

	Ag	Al	Be	Bi	Cd	Co	Cr	Cu	Fe	Ga	Mg	Mn	Na
1	0.0002	84.010	0.0002	0.0220	0.0010	0.0048	0.0721	3.2900	0.5580	0.0096	0.1000	0.1900	0.0002
2	0.0002	84.120	0.0002	0.0222	0.0010	0.0050	0.0721	3.3100	0.5700	0.0100	0.1020	0.1922	0.0006
3	0.0005	84.500	0.0003	0.0230	0.0013	0.0050	0.0731	3.3690	0.5708	0.0104	0.1021	0.1950	0.0082
4	0.0006	84.600	0.0003	0.0230	0.0013	0.0050	0.0820	3.3700	0.5790	0.0108	0.1050	0.1960	0.0756
5	0.0006	84.600	0.0003	0.0230	0.0015	0.0050	0.0830	3.3990	0.5830	0.0113	0.1210	0.1960	<0.001
6	0.0007	84.692	0.0003	0.0240	0.0016	0.0051	0.0858	3.4000	0.5850	0.0130	0.1214	0.1960	
7	<0.001		0.0003	0.0256	0.0016	0.0055	0.0860	3.4064	0.5860	0.0130	0.1220	0.1960	
8	<0.001		0.0004	0.0260	0.0020	0.0057	0.0865	3.4140	0.6000	0.0130	0.1240	0.1980	
9	<0.002		<0.0002	0.0269	0.0020	0.0057	0.0866	3.4530	0.6027	0.0131	0.1260	0.1999	
10	<0.002		<0.001	0.0275	0.0023	0.0057	0.0870	3.4580	0.6040	0.0134	0.1267	0.2019	
11	<0.002			0.0279	0.0025	0.0060	0.0872	3.4600	0.6070	0.0143	0.1290	0.2020	
12				0.0286	0.0026	0.0063	0.0890	3.4600	0.6071	0.0163	0.1300	0.2024	
13				0.0291	<0.002	<0.01	0.0894	3.4640	0.6080	0.0167	0.1300	0.2040	
14				0.0306	<0.002	<0.01	0.0900	3.4870	0.6160		0.1340	0.2050	
15					<0.002		0.0910	3.5080	0.6200		0.1350	0.2070	
16					<0.01		0.0927	3.5080	0.6247		0.1357	0.2080	
17							0.0937	3.5759	0.6270		0.1359	0.2090	
18							0.0950	3.5800	0.6450		0.1390		
19							0.1010		0.6500				
Mean	0.0005	84.420	0.0003	0.0257	0.0017	0.0054	0.0865	3.4396	0.6023	0.0127	0.1233	0.1999	0.0212
STDV	0.0002	0.2840	0.0001	0.0028	0.0006	0.0005	0.0076	0.0789	0.0250	0.0022	0.0127	0.0056	0.0365
Certified	0.0005	(84.4)	0.00029	0.026	0.0017	0.0054	0.086	3.44	0.60	0.013	0.12	0.200	(0.02)
U_{CRM}	0.0002		0.00006	0.002	0.0004	0.0003	0.004	0.04	0.01	0.001	0.01	0.003	
Methods	I,O,IM,X	I,O,X	I,O,IM	I,O,X,IM	I,IM,O,G,X	I,IM,O,X,G	I,O,X,G,IM	I,O,X,G	I,O,X,G	I,IM,O,X,G	I,O,X,G,IM	I,O,X,IM	I

	Ni	P	Pb	S	Sb	Si	Sn	Sr	Ti	V	Zn	Zr
1	0.5000	0.0009	0.1100	0.0001	0.0428	7.7240	0.2330	0.0001	0.0870	0.0157	2.0500	0.0005
2	0.5100	0.0010	0.1126	0.0011	0.0430	7.7300	0.2360	0.0002	0.0930	0.0157	2.0700	0.0007
3	0.5120	0.0015	0.1200	0.0049	0.0480	7.7700	0.2430	0.0004	0.0968	0.0160	2.0900	0.0007
4	0.5148	0.0019	0.1230	0.0053	0.0480	7.8300	0.2502	0.0030	0.0980	0.0160	2.0975	0.0008
5	0.5170	0.0023	0.1257	0.0087	0.0480	7.9411	0.2510	<0.0001	0.1000	0.0166	2.1000	0.0009
6	0.5200	0.0030	0.1270	0.0100	0.0500	8.0250	0.2610	<0.0001	0.1000	0.0166	2.1170	0.0010
7	0.5220	0.0040	0.1310	0.0100	0.0500	8.0340	0.2630	<0.0002	0.1019	0.0169	2.1300	0.0012
8	0.5220	0.0047	0.1317	<0.0001	0.0509	8.0404	0.2650	<0.0005	0.1020	0.0170	2.1400	0.0014
9	0.5227	0.0050	0.1327		0.0518	8.0700	0.2700	<0.0005	0.1040	0.0170	2.1460	0.0015
10	0.5230	<0.005	0.1330		0.0547	8.1200	0.2756	<0.0005	0.1050	0.0172	2.1570	<0.0005
11	0.5290	<0.005	0.1380		0.0556	8.1255	0.2790	<0.001	0.1070	0.0173	2.1820	<0.0005
12	0.5300	<0.005	0.1393		0.0556	8.1305	0.2849		0.1072	0.0174	2.1866	<0.001
13	0.5310		0.1400		0.0606	8.1800	0.2850		0.1073	0.0175	2.1900	<0.001
14	0.5320		0.1400			8.1940	0.2850		0.1085	0.0180	2.2094	
15	0.5332		0.1410			8.2350	0.2914		0.1090	0.0180	2.2263	
16	0.5392		0.1430			8.3300	0.2950		0.1100	0.0185	2.2460	
17	0.5392		0.1439			8.5293	0.3000		0.1100	0.0185	2.2890	
18			0.1470				0.3009		0.1140		2.3540	
19			0.1606				0.3393		0.1250			
Mean	0.5234	0.0027	0.1337	0.0057	0.0507	8.0593	0.2741	0.0009	0.1045	0.0171	2.1656	0.0010
STDV	0.0106	0.0016	0.0123	0.0041	0.0051	0.2157	0.0262	0.0014	0.0082	0.0009	0.0790	0.0003
Certified	0.523	0.003	0.134	(0.006)	0.051	8.1	0.27	(0.001)	0.105	0.0171	2.17	0.0010
U_{CRM}	0.005	0.001	0.006		0.003	0.1	0.01		0.004	0.0005	0.04	0.0003
Methods	I,O,X,G	I,O,IM,X,W	I,O,X,G,IM	I,C,X	I,X,IM,O	I,O,W,G,X	I,O,X,G,IM	I,O,IM,X	I,O,X,G,IM	I,O,X,G	I,IM,O,X	

Legend: W = Classical, C = Combustion, F = Fusion, A = AA or GFAA, I = ICP or DCP, IM=ICP-MS, D = DC Arc, O = AES, X = XRF, G = GDAES or GDMS, H = Hollow Cathode AES